


# Vita<sup>OKmeter</sup>

## Blood Glucose Monitoring System



strip Ejection



Results in  
6 seconds



Only 0.7  $\mu$ L  
of blood sample

- Accurate results in 6 seconds.
- Small blood sample, only 0.7  $\mu$ L.
- 1000 test memory with date and time.

# SPECIFICATIONS

Model Name	OK-C
Assay Method	Electrochemical biosensor
Test Sample	Capillary whole blood
Test Result	Referenced to plasma glucose value
Alternative Site Testing	YES (palm, forearm, upper arm, calf, or thigh)
Sample Size	0.7 µL
Measuring Time	6 seconds
Measuring Range	20~600 mg/dL
Acceptable Hematocrit Range	20~60%
Operating Condition	10°C~40°C (50°F~104°F), between 10~85% R. H.
Storage/Transportation Condition	4°C~40°C (39°F~104°F), between 10~85% R. H.
Memory Capacity	1000 test results with time and date
Average Calculation	7, 14 and 28 days
Power Supply	CR2032 (3V) battery x 1
Battery Life	Approximately 1,000 tests
Dimensions	79 x 45 x 20 mm
Weight	76 g (with batteries)
Pollution Degree 2	In the intended environment



Manufacturer

**OK Biotech Co., Ltd.**

No.91, Sec. 2, Gongdao 5th Rd., Hsinchu City 30070, Taiwan

[www.okbiotech.com](http://www.okbiotech.com)